



America

CERTIFICATE

No. U8 17 03 64872 070

Holder of Certificate: Sheldon Manufacturing, Inc.

300 N. 26th Avenue
Cornelius OR 97113
USA

Certification Mark:



Product: Laboratory Equipment
(HF Ovens)

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 231-72105089-100

Valid until: 2022-03-16

Date, 2017-04-21

Page 1 of 2





America

CERTIFICATE

No. U8 17 03 64872 070

Model(s):

1675, 1680, 1685, 1690, CR1-2, HF4-2, HF10-2, HF15-2,
 HF25-2, HF37-2, SMO5CR-2, SMO5HP-2, SMO10HP-2, SMO14HP-2,
 SMO28HP-2, SMO34HP-2, SMO38HP-2, TFO28, TFO10, FH5, TFO-5CR,
 TFO-5N, TFO-14, TFO-38, CDO-10, CDO-14, CDO-28, CDO-38

Brand Name:

Shellab®, Cascade TEK, TestEQUITY, VWR, Cascade Sciences

Parameters:

Rated Input Voltage: 220-240VAC
 Rated Frequency: 50/60Hz
 Rated Input Power:
 12.0A:CR1-2, HF4-2, SMO5HP-2, SMO5CR-2, FH5, TFO-5CR, TFO-5N
 26.0A:1675, 1680, HF10-2, HF15-2, SMO10HP-2, SMO14HP-2,
 TFO10, FH5, TFO-14, CDO-10, CDO-14, CDO-28, CDO-38
 50.0A:1685, 1690, HF25-2, HF37-2, SMO28HP-2, SMO34HP-2,
 SMO38HP-2, TFO28, TFO-38
 Protection Class: PE Connected
 Ingress Protection: IPX0
 Ambient Temperature: 40°C

Conditions of Acceptability:
 These units are not intended for use
 at hazardous or household locations.

**Tested
according to:**

CAN/CSA-C22.2 No. 61010-1:2012
 CAN/CSA-C22.2 No. 61010-2-010:2015
 UL 61010-1:2012
 UL 61010-2-010:2015
 EN 61010-1:2010
 EN 61010-2-010:2014

**Production
Facility(ies):**

64872